

4

3

2

1

REVISIONS				
ZONE	REV	DESCRIPTION	DATE	APPROVED
	1.0	Initial Release	03/22/04	LBK
	2.0	Redesign P-trap	02/18/05	LBK

D

D

C

C

B

B

A

A

ELECTRONICS UNIT WITH  
CONTROLS AND DIGITAL READOUT

0.42



10 PSI  
(.7 BAR)

1/4" NPT

Note: Install shutoff  
valve between DOM-1  
Stripper Assembly &  
process stream

3/8" (9-10mm) ID flexible plastic tubing

ONLY AIR AND AIR/OZONE  
SHOULD FLOW THROUGH  
THESE TUBES

KEEP EXIT TUBING  
BELOW STRIPPER SO THAT  
WATER FLOWING IN TUBING  
DOES NOT BACK UP INTO STRIPPER

ECO SENSORS, INC.  
Santa Fe, NM  
505-988-1421  
larry@ecosensors.com

DOM-1 SYSTEM PLACEMENT  
Drawing

Designed by: Larry Kilham	SIZE	FSCM NO.	DWG NO.	Designed by: Larry Kilham	REV
Checked by: Larry Kilham	SCALE	N/A	Drawn by: Ed Tomingas	SHEET	N/A

4

3

2

1