


OZONE GENERATOR PERFORMANCE TEST

OZV-8 Ozone Generator


#1

	Oxygen Flow in LPM	Ozone Concentration, % by weight	Ozone Production in g/hr	PSI	COMMENTS
	1.00	4.77	4.09	3	
	2.00	3.70	6.35	3	
	3.00	2.95	7.59	3	
	4.00	2.42	8.31	3	
	5.00	2.07	8.88	3	
	6.00	1.73	8.91	3	
	7.00	1.50	9.01	3	
	8.00	1.30	8.92	3	

Ozone Concentration



Ozone Production



Test was performed at 72°F, Generator was ON for more than 1 hour.

Ozone Analyzer API-460H & API-454